

CALIBRATION CERTIFICATE

Total Weighing Solutions Ltd
Unit 8 Signal Park, West March
Daventry, Northants, NN11 4RU, UK
www.totalweighing.co.uk
Tel: 01327 368020



Customer Details

Name: WHILTON MILL LTD
Address: WHILTON LOCKS

DAVENTRY
NORTHAMTONSHIRE
NN1 2NH
Contact: DAN
Telephone: 01327 843822

Certificate No. 19283

Date of Calibration: 29/11/2024
Next Due: 29/11/2025

Location:
Department:
Manufacturer: DINI ARGEO
Model: DFWLB-2
Serial No.: 0722941484
Capacity: 600 Kg
Resolution: 0.1
Tolerance: +/- 3 divs
Calibration Frequency: Yearly

Test Weight/Rig Details

Identification No. and Certificate No.

WEIGHT SET:
T1 - T35, D11724 & B12267


CERTIFICATE NUMBER:
12582, 11301, 11303 & 121466

Test weights and equipment are traceable to National Standards

Linearity / Accuracy Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0	0.0	0.0	0.0	0.0
60	59.9	-0.1	60.0	0.0
120	119.8	-0.2	120.0	0.0
200	199.7	-0.3	200.0	0.0
300	299.6	-0.4	300.0	0.0

Diagram



Off Centre Load Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0

Repeatability Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
200.0	200.0	0.0	200.0	0.0
200.0	200.0	0.0	200.0	0.0
200.0	200.0	0.0	200.0	0.0

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments

Comments

Tolerances

Unless otherwise agreed, tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:-
0-500 divisions +/- 1div
>501-2000 divisions +/- 2div
>2001 divisions +/- 3div

Technicians Name

Tony Goodman

Technicians Signature

Total Weighing Solutions Ltd

Registered Office: Unit 8 Signal Park, West March, Daventry, Northants, NN11 4RU, UK

Posting Time : 21:02