



CALIBRATION CERTIFICATE

Total Weighing Solutions Ltd
 Unit 8 Signal Park
 Daventry
 Northamptonshire
 NN11 4RU
 Tel: 01327 368020
 www.totalweighing.co.uk



Customer Details

Name: WHILTON MILL LTD
 Address: WHILTON LOCKS

 DAVENTRY
 NORTHAMTONSHIRE
 NN1 2NH
 Contact: DAN
 Telephone: 01327 843822

Linearity / Accuracy Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
0.0	0.0	0.0	0.0	0.0
60.0	60.0	0.0	60.0	0.0
120.0	120.1	0.1	120.0	0.0
200.0	200.2	0.2	200.0	0.0
300.0	300.4	0.4	300.0	0.0

Certificate No. 20676

Date of Calibration: 26/11/252025
Next Due: 26/11/262026
Sales Order:

Diagram



Scale Details

Location:
 Manufacturer: DINI ARGEO
 Model No.: DFWLB-2
 Serial No.: 722941484
 Capacity: 600 Kg
 Resolution: 0.1
 Tolerance: +/- 3 divs
 Frequency: Yearly

Off Centre Load Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0
100.0	100.0	0.0	100.0	0.0

Test Weight/Rig Details

Identification No. and Certificate No.
 Test Weights and Equipment are traceable to National Standards

WEIGHT SET:
 T1 - T35, D11724 & B12267

CERTIFICATE NUMBER:
 13083, 11470, 11471 & 121466

Repeatability Test

Test Point (Kg)	As Found Reading	As Found Deviation	Definitive Reading	Definitive Deviation
300.0	300.0	0.0	300.0	0.0
300.0	300.0	0.0	300.0	0.0
300.0	300.0	0.0	300.0	0.0

This certificate complies with the United Kingdom Weighing Federation Calibration Code of Practice for non-automatic weighing instruments

Comments

Tolerances

Unless otherwise agreed, tolerances will be those permitted by current Weights and Measures Regulations for Class III machines as follows:-
 0-500 divisions +/- 1div
 >501-2000 divisions +/- 2div
 >2001 divisions +/- 3div

Technicians Name: Tony Goodman

Technician Signature:

Total Weighing Solutions Ltd
 Registered Office: Unit 8 Signal Park, Daventry, Northamptonshire, NN11 4RU, UK

Posting Time : **16:00**