

TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

Customer Cumbria Kart Racing Club				Customer Order No.	
				Date of Last Test	1 Feb 2025
				Date of Test	21 Feb 2026
				Next Test Due	21 Feb 2027
Site Address :- Rowrah Cumbria				Next Test Due within 30 days of above.	
				Machine	Platform
				Type	
				Manufacturer	TopWeigh Ltd
				Capacity :- Max	300 kg
				Min	
				e.	0.1 kg
Calibration Certificate No.		TW	1026111	Maximum Tolerance	
Initial Test		Calibration Test		Serial No.	
Weights Applied	Reading	Weights Applied	Reading		
				Class	(111)
				Observations & Comments :-	
				Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT	
				Weights Traceable to National Standards TW 11985	
				Weights Used TWR Set	


This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :-  Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.

Posting Time : **15:15**



TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

Customer Cumbria Kart Racing Club				Customer Order No.	
				Date of Last Test	1 Feb 2025
				Date of Test	21 Feb 2026
				Next Test Due	21 Feb 2027
Site Address :- Rowrah Cumbria				Next Test Due within 30 days of above.	
				Machine	Dail Platform
				Type	Avery
				Manufacturer	4208 GLE
				Capacity :- Max	250 kg
				Min e.	500g
Calibration Certificate No.		TW	1026112	Maximum Tolerance	
Initial Test		Calibration Test		Serial No.	
Weights Applied	Reading	Weights Applied	Reading		
				Observations & Comments :-	
				Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT	
				Weights Traceable to National Standards TW 11985	
				Weights Used TWR Set	

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :- Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.

Posting Time :

15:15



TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile : +44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

Customer Cumbria Kart Racing Club				Customer Order No.	
				Date of Last Test	1 Feb 2025
				Date of Test	21 Feb 2026
				Next Test Due	21 Feb 2027
Site Address :- Rowrah Cumbria				Next Test Due within 30 days of above.	
				Machine	
				Type	Test Weights
				Manufacturer	
				Capacity :-	Max
					Min
Calibration Certificate No.				TW	1026113
Initial Test				Calibration Test	
Weights Applied	Reading	Weights Applied	Reading	Serial No.	
				RCC2-RCC6	
RCC 2 20 Kg RCC 3 20 Kg RCC 4 20 Kg RCC 5 20 Kg RCC 6 20 Kg				Observations & Comments :-	
				Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016	
				Found to be CORRECT	
				Weights Traceable to National Standards	
				TW 11985	
				Weights Used TWR Set	


This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :-  Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.

Posting Time : **15:15**



TopWeigh Ltd

Scale House, (Jeffrey Estates), Rockcliffe,
Carlisle, Cumbria. CA6 4BH. UK



Tel : +44(0)1228 672400, Fax : +44(0)7092 884848
Mobile :+44 (0)7808 308689, ajj@scalemaker.co.uk

Calibration Certificate

Customer Cumbria Kart Racing Club				Customer Order No.	
				Date of Last Test	
				Date of Test	26 Mar 2026
				Next Test Due	26 Mar 2027
Site Address :- Rowrah Cumbria				Next Test Due within 30 days of above.	
				Machine	Platform
				Type	
				Manufacturer	TopWeigh Ltd
				Capacity :- Max	300 kg
				Min e.	0.1 kg
Calibration Certificate No.		TW	1026117	Maximum Tolerance	
Initial Test		Calibration Test		Serial No.	0742720666
Weights Applied	Reading	Weights Applied	Reading	Class	(111)
				Observations & Comments :-	
				Serviced and then Tested for compliance within Inspection tolerances of Non Automatic weighing Instruments regulations 2016 Found to be CORRECT	
				Weights Traceable to National Standards TW 11985	
				Weights Used TWR Set	

This is to certify that the equipment listed was tested in accordance with EN45 501-8.2

Engineers Sig :-  Print Name :- A J Jeffrey

Weighing Solutions on a Global Scale

Mechanical. Electronic. Hydraulic. Pneumatics. Calibration. Software. Design. Manufacture. Service. Repair.

Posting Time : **15:15**